

1. The demand for full-arch implant restorations has increased due to the rise in population lifespan and advancements in digital technologies.
 - a. True
 - b. False
2. Multi-unit abutments (MUAs) are primarily used to eliminate the need for impressions during full arch restorations.
 - a. True
 - b. False
3. The screw-titanium connection relies on cement to attach the titanium base to the zirconia prosthesis.
 - a. True
 - b. False
4. The direct screw-zirconia connection eliminates the cement layer, reducing potential biological complications.
 - a. True
 - b. False
5. Conventional prosthetic screws designed for titanium frameworks work equally well for direct zirconia-MUA connections.
 - a. True
 - b. False
6. The Vortex screw can be used in both screw-titanium and screw-zirconia connections.
 - a. True
 - b. False
7. The Vortex screw is compatible with angulated screw channels up to 30 degrees without compromising torque delivery.
 - a. True
 - b. False
8. Applying more than 20 N.cm of torque to the Vortex screw will fracture the screw itself.
 - a. True
 - b. False
9. The Vortex screw system is compatible with multiple implant platforms and integrates with CAD software such as exocad and 3Shape.
 - a. True
 - b. False
10. The primary purpose of the Vortex screw is to increase the number of prosthetic components used in full-arch restorations.
 - a. True
 - b. False

Passing quiz grades are worth ½ point documented scientific credit. To earn CDT credit, once the quiz is completed, send it to the NADL at the address or fax number below or submit this quiz online at <https://www.nadl.org/jdt-quiz-submissions>. To earn an additional ½ point professional development credit, visit www.nbccert.org to submit your time for reading the accompanying article(s) in the professional development log. Quiz credits will appear on the NBC CDT Online Education Tracking System at www.nbccert.org, which is updated weekly. This quiz is provided to test the technician's comprehension of the article's content, and does not necessarily serve as an endorsement of the content by NADL or NBC.

Name: _____ CDT #: _____ Date: _____